

OFFICE OF THE REGULATED MARKET COMMITTEE, BALASORE

AT-BAMAPADA PO-CHHANPUR VIA-KURUDA DIST-BALASORE PIN-756056

Letter No. 550 /Date- 29.10.2021**Quotation Call Notice**

Sealed Quotations are invited from manufacturers / Suppliers having valid GST number / Firm Registration Certificate / Authorization Certificate to supply the following Grading Equipments to the Regulated Market Committee, Balasore & Jaleswar for Paddy Procurement Operations during KMS 2021-22. The last date for submission of quotation is dt. 11.11.2021 till 11.30 A.M. and the same will be opened on dt. 11.11.2021 at 2.00 PM in the presence of the quotationers or their authorized representatives in the office chamber of the Sub-Collector, Balasore-cum-Chairman, Regulated Market Committee, Balasore & Jaleswar & the date of demonstration of the equipments will be on date-12.11.2021 at 11.00 AM. The quotations must be sent through Registered Post / Speed Post only in the Address of "**The Secretary, Regulated Market Committee, Balasore, At : Bamapada, P.O.: Chhanpur, Via : Kuruda, Dist.: Balasore, Pin : 756056**". The authority is not responsible for any postal delay. The quotations shall be finalized by a Tender Evaluation Committee and the purchase order shall be placed after physical verification of the equipments. The price must include all taxes, transportation and other cost etc..


The authority reserves the right to reject any or all quotations at any time without assigning any reason thereof.

Specifications of the Equipments as per O.S.A.M. Board, Bhubaneswar Letter No.3252(66) dated-30.10.2018 are as follows.

Sl. No.	Name of the Item	Specification of the Equipments	Rate per Unit
1	Electronic Weighing Scale of 300 Kg Capacity (Without Printer)	<p>Minimum Weight of the Machine : 30 Kg.</p> <p>Weighing Capacity : 300 Kg. maximum, 1Kg minimum, e=50g (least count).</p> <p>Accuracy Classes : Class-III.</p> <p>Platform Size : 60cm x 60 cm.</p> <p>Platform : Powder coated frame and Stainless Steel Platter (202 Grade & above).</p> <p>Indicator / Display : LED / LCD Display Screen</p> <p>Load cell : The load cell must be overloaded and shock protection, connecting wires from base to head should be of rat protection.</p> <p>Battery : Minimum 12 hours & above battery backup, battery housing should be separated and should be replaces without breaking seal of Inspector, Legal Metrology.</p> <p>Over load : 600 Kg & above.</p> <p>Model Approval : Weighing Scale must have Model Approval Certificate from competent authority of Govt. of India.</p> <p>Machines shall be delivered after Certification by the LMO.</p>	

Contd...P/02

2	Analysis Kit Box	Analysis Box / Kit consists of Class 'B' Mechanical Weighing Balance up to 100gm., Weight box consist of set of verified weights up to 100gms. Enameled Plate-9", Parkhi-9", Sieve-3.30 mesh, Magnifying Glass-1.5"(Magnification of 10X), Vernier / Micrometer, Palm Paddy De-husker, Scoop (With handle, 105mm x 1000mm x 25mm, Measuring Cylinder-20cc., Built-in Purity Work Board for testing physical purity of seeds, Petri Dish & Forceps all packed in a compact wooden/ leather brief case, as small, as light and as compact as possible, but as accurate and as reliable.	
3	Seed Sample Divider (Boerner type)	i. Meeting BIS specification ii. No. of Pockets : 38 iii. Hopper Capacity: 1Kq. (minimum) iv. Feature: provides accurate splits of the grain Samples into two equal halves with an accuracy of +/-5% on a 1Kg sample. v. Two collecting Pans vi. Copper and Brass Body vii. Made without joint (in material flowing component).	
4	Set of Sieves	Set of Sieves (4 nos. with collecting Pan) and lid made up of brass, Set of Sieves Size - 4mm : 01 No., 3.5mm : 01 No., 1.7mm : 01 No., 1mm : 01 No., Dust Collector : 01 No., Solid Bottom Pan : 01 No., Lid : 01 No.	
5	Digital Moisture Meter	Measuring Range : 3.5% to 40% Supply Voltage : 9V DC Temperature Compensation : Automatic Calibration & error correction facility, Display : Digital LCD / LED Error Condition Display for Over range or under range : $\pm 0.2\%$ Commodity : Cereals, Pulses, Oil Seeds. Operating Temperature : Ambient to 50° C	
6	Parkhee	Size : 12" (Stainless Steel)	


Sub-Collector, Balasore
-Cum-Chairman, R.M.C., Balasore
R

09.10.21